

Certificate of Analysis



Cathode Buffer Container 3500/Flex Series

Product No. 4408256
Lot No. 2505090
Date of Manufacture 19MAY2025
Expiration Date 17JUL2026

TEST	SPECIFICATION	RESULT
Visual Inspection		
Clear, particulate free, viscous solution		Pass
pH Test		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 – 8.2	7.9
Conductivity		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 – 2000 µS/cm	1802 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.
Site: Warrington, UK

Quality Assurance
Issued 13JUN2025

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

CRT1000874_A