

# Certificate of Analysis



## Anode Buffer Container 3500/Flex Series

Product No. 4393927  
Lot No. 2502007  
Date of Manufacture 10MAR2025  
Expiration Date 30APR2026

TEST	SPECIFICATION	RESULT
<b>Physical Inspection</b>		
Material is clear, colourless and free of particulates		
Appearance - Conformance		Pass
<b>pH Test</b>		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 – 8.2	7.7
<b>Conductivity</b>		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 – 2000 µS/cm	1789 µS/cm

**For Research Use Only. Not for use in diagnostic procedures.**

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.  
Site: Warrington, UK

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
**Quality Assurance**  
**Issued 17APR2025**

CRT1000871\_A