

Certificate of Analysis



Anode Buffer Container 3500/Flex Series

Product No. 4393927
Lot No. 2505028
Date of Manufacture 09JUN2025
Expiration Date 04AUG2026

TEST	SPECIFICATION	RESULT
Physical Inspection		
Material is clear, colourless and free of particulates		
Appearance - Conformance		Pass
pH Test		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 – 8.2	7.9
Conductivity		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 – 2000 µS/cm	1827 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.
Site: Warrington, UK

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

Quality Assurance
Issued 13JUN2025

CRT1000871_A