

Certificate of Analysis



SeqStudio™ Genetic Analyzer Cathode Buffer Container

Product No. A33401
Lot No. 2503168
Date of Manufacture 25APR2025
Expiration Date 01MAY2026

TEST	SPECIFICATION	RESULT
Physical Inspection		
Material is clear, colourless and free of particulates		
Appearance - Conformance		Pass
pH Test		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 – 8.2	7.8
Conductivity		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 – 2000 µS/cm	1741 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.
Site: Warrington, UK

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com


Quality Assurance
Issued 20JUN2025

CRT1000878_A