

# Certificate of Analysis



## SeqStudio™ Genetic Analyzer Cathode Buffer Container

Product No. A33401  
Lot No. 2506178  
Date of Manufacture 13JUL2025  
Expiration Date 02OCT2026

TEST	SPECIFICATION	RESULT
<b>Physical Inspection</b>		
Material is clear, colourless and free of particulates		
Appearance - Conformance		Pass
<b>pH Test</b>		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 – 8.2	7.9
<b>Conductivity</b>		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 – 2000 µS/cm	1807µS/cm

**For Research Use Only. Not for use in diagnostic procedures.**

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.  
Site: Warrington, UK

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
Quality Assurance  
Issued 23JUL2025

CRT1000878\_A