

Certificate of Analysis



Cathode Buffer Container 3500/Flex Series

Product No. 4408256
Lot No. 2509114
Date of Manufacture 03OCT2025
Expiration Date 24DEC2026

TEST	SPECIFICATION	RESULT
Visual Inspection		
Clear, particulate free, viscous solution		Pass
pH Test		
Measured at 25.0°C ± 0.3°C		
pH Value	7.6 – 8.2	8.0
Conductivity		
Measured at 25.0°C ± 0.1°C		
Conductivity	1500 – 2000 µS/cm	1692 µS/cm

For Research Use Only. Not for use in diagnostic procedures.

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.
Site: Warrington, UK

A handwritten signature in black ink, appearing to be a stylized 'L' or similar character.

Quality Assurance
Issued 08OCT2025

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com

CRT1000874_A