

Certificate of analysis

ProcartaPlex

 Cat. number: **EPX010-26079-901**

96 tests/1 analyte

 Name: **ProcartaPlex Mouse BLC
(CXCL13) Simplex**

 Lot number: **419543-004**

 Expiry Date: **2026-10**

Components

Components	Quantity	Lot	Store at
S20837EX Standardmix H	2 each	402510-000	2-8°C
BD26079EX BLC Mag Beads	1 x 120µl (50x)	411634-000	2-8°C
BK26079EX BLC Det. Ab.	1 x 70µl (50x)	411635-000	2-8°C

Standardmix H Lot#402510-000

Provided below is a table of Standard 1 (Std1) value for each analyte in each tube when prepared according to the "Prepare antigen standard" procedure of the Manual. All data were evaluated in a 2-hour incubation at room temperature.

Analyte	Std1 Concentration (pg/ml)	ULOQ / LLOQ (pg/ml)
		Determined in Universal Assay Buffer
BLC (CXCL13)	8300	8300 / 2,03
CTACK (CCL27)	33700	8425 / 8,23
Eotaxin-2 (CCL24)	3550	3550 / 0,87
Granzyme B	23700	23700 / 5,79
IL-6R/sIL-6R	2462400	2462400 / 601
I-TAC (CXCL11)	10600	2650 / 2,59
TARC (CCL17)	3550	3550 / 0,87
TECK (CCL25)	514400	128600 / 126

BLC Mag Beads Lot#411634-000

Target Name	Bead Number	Concentration Std1	Standard
BLC (CXCL13)	56	8300 pg/ml	Standardmix H

Analytical information:

This product has been tested by Quality Control and passed internal specifications.

Quality control:

