

# Certificate of Analysis

## ViralSEQ™ dPCR Lentivirus Physical Titer Kit

**Product No.** A59317  
**Lot No.** 2505003  
**Date of Manufacture** 23MAY2025  
**Expiration Date** 05MAR2026

### KIT CONFIGURATIONS:

COMPONENT DESCRIPTION	LOT NO.
100122791 Tube, Absolute Q RT-dPCR Master Mix	3239391
100122788 Tube, dPCR Lentivirus Physical Titer Assay Mix	2503003
100102063 Tube, Nuclease Free Water	2504018

TEST	SPECIFICATIONS	RESULT
------	----------------	--------

QC Method: The positive results obtained when analyzed on the QuantStudio Absolute Q Digital PCR System are the following:

Lentiviral RNA (FAM)	Pass Functional Test	Pass
IPC (VIC)	90 – 220 cp/µL	Pass

QC Method: The negative results obtained when analyzed on the QuantStudio Absolute Q Digital PCR System are the following:

No Template Control (NTC)	Pass Functional Test	Pass
---------------------------	----------------------	------

*For Research Use Only. Not for use in diagnostic procedures.*

Manufactured in compliance with our ISO 13485 certified quality management system.

**Site:** Warrington, UK



Quality Assurance  
Issued 17JUN2025

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)