

# Certificate of Analysis

resDNASEQ™ dPCR CHO DNA Kit

Product No. A59366  
Lot No. 2503003  
Date of Manufacture 11APR2025  
Expiration Date 06AUG2026

## KIT CONFIGURATIONS:

COMPONENT DESCRIPTION	LOT NO.
A52490 Tube, Absolute Q DNA Digital PCR Master Mix (5X)	3193162
100122936 Tube, dPCR CHO DNA Assay Mix	2411003
100102063 Tube, Nuclease Free Water	2410017

TEST	SPECIFICATIONS	RESULT
------	----------------	--------

QC Method: The positive results obtained when analyzed on the QuantStudio Absolute Q Digital PCR System are the following:

CHO DNA (FAM)	Pass Functional Test	Pass
IPC (VIC)	45 – 115 cp/μL	Pass

QC Method: The negative results obtained when analyzed on the QuantStudio Absolute Q Digital PCR System are the following:

No Template Control (NTC)	Pass Functional Test	Pass
---------------------------	----------------------	------

*For Research Use Only. Not for use in diagnostic procedures.*

Manufactured in compliance with our ISO 13485 certified quality management system.

Site: Warrington, UK

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
Quality Assurance  
Issued 13MAY2025