

# Certificate of Analysis



## GeneScan™-500 LIZ™ Size Standard

Product No. **4322682**  
Lot No. **2411487**  
Date of Manufacture **19DEC2024**

### KIT CONFIGURATION

Part No.	Component Description	Lot No.
4322679	Tube, GS500 Liz Size Standard	2409197

TEST	SPECIFICATION	RESULT
------	---------------	--------

#### Physical Inspection

Tubes are inspected for correct label and lot number. Caps are tight with no signs of leakage.

PASS

PASS

### Use Test

Each lot of GeneScan™-500 LIZ™ Size Standard is tested for correct fragment pattern and peak height on a 3500xl Genetic Analyzer.


PASS

PASS

***For Research Use Only. Not for use in diagnostic procedures.***

Manufactured in compliance with our ISO 13485 certified quality management system. Site: Warrington, UK

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at [cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com).

  
Quality Assurance  
Issued 03JAN2025