

## TAQMAN ENDOGENOUS CONTROLS FG,HUMAN GAPD MGB

Product No. 4352934E  
Lot No. 2502098  
Date of Manufacture 09MAR2025

TEST	RESULT
<b>ANALYTICAL TESTS</b>	
Mass Spectrometry specification + /- 0.3%	
Mass Spectrometry (Target Calculated Mass)	Pass
Minimum Yield Specification	Pass
<b>FUNCTIONAL TEST</b>	
Test Material is within + /- 1CT of reference lot	
Real Time PCR functional test	Pass

*For Research Use Only.*  
*Not for use in diagnostic procedures.*

**Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001. Site: Warrington, UK**



**Quality Assurance**  
**Issued 11MAR2025**

Corporate Legal Entity  
Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)