

TAQMAN ENDOGENOUS CONTROLS FG,HUMAN GAPD MGB


Product No. 4352934E
Lot No. 2504099
Date of Manufacture 16APR2025

TEST	RESULT
ANALYTICAL TESTS	
Mass Spectrometry specification + /- 0.3%	
Mass Spectrometry (Target Calculated Mass)	Pass
Minimum Yield Specification	Pass
FUNCTIONAL TEST	
Test Material is within + /- 1CT of reference lot	
Real Time PCR functional test	Pass

For Research Use Only.
Not for use in diagnostic procedures.

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001. Site: Warrington, UK

Corporate Legal Entity
Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com


Quality Assurance
Issued 17APR2025