

# Certificate of Analysis



## Hi-Di™ Formamide – 5mL


Product No. 4401457  
Lot No. 2503198  
Date of Manufacture 08APR2025  
Expiration Date 30MAY2026

TEST	SPECIFICATION	RESULT
<b>Conductivity</b>		
Measured at 23.0°C ± 0.1°C		
Conductivity	19 – 34 µS/cm	29 µS/cm

**For Research Use Only. Not for use in diagnostic procedures.**

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.  
Site: Warrington, UK

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
**Quality Assurance**  
Issued 25APR2025

CRT1000872\_A