

# Certificate of Analysis



## Hi-Di™ Formamide – 5mL

Product No. **4401457**  
Lot No. **2505200**  
Date of Manufacture **13MAY2025**  
Expiration Date **30MAY2026**

TEST	SPECIFICATION	RESULT
<b>Conductivity</b>		
Measured at 23.0°C ± 0.1°C		
Conductivity	19 – 34 µS/cm	29 µS/cm

**For Research Use Only. Not for use in diagnostic procedures.**

Manufactured under a Quality Management System that meets the requirements of ISO 13485 and ISO 9001.  
Site: Warrington, UK

Life Technologies  
5781 Van Allen Way  
Carlsbad, CA, USA 92008  
[www.thermofisher.com](http://www.thermofisher.com)  
For inquiries, contact us at  
[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

  
**Quality Assurance**  
**Issued 20JUN2025**

CRT1000872\_A