Bender MedSystems GmbH Campus Vienna Biocenter 2 A-1030 Vienna, Austria www.thermofisher.com

Certificate of analysis

ProcartaPlex

Cat. number:	EPX010-26074-901 96 tests/1 analyte	Name:	ProcartaPlex Mouse Granzyme B Simplex		
Lot number:	414994-003	Expiry Date:	2026-11		
Components S20837EX	Standardmix H		Quantity Lot Store at 2 each 441372-000 2-8°C		

S20837EX	Standardmix H	2 each	441372-000	2-8°C
BD26074EX	Granzyme B Mag Beads	1 x 120µl (50x)	402913-000	2-8°C
BK26074EX	Granzyme B Det. Ab.	1 x 70µl (50x)	402914-000	2-8°C

Standardmix H Lot#441372-000

Provided below is a table of Standard 1 (Std1) value for each analyte in each tube when prepared according to the "Prepare antigen standard" procedure of the Manual. All data were evaluated in a 2-hour incubation at room temperature.

Analyte	Std1 Concentration (pg/ml)	ULOQ / LLOQ (pg/ml) Determined in Universal Assay Buffer
BLC (CXCL13)	6650	6650 / 1,62
CTACK (CCL27)	17000	17000 / 4,15
Eotaxin-2 (CCL24)	2300	2300 / 2,25
Granzyme B	21600	21600 / 5,27
IL-6R/sIL-6R	1789700	1789700 / 437
I-TAC (CXCL11)	7400	7400 / 1,81
TARC (CCL17)	3350	3350 / 0,82
TECK (CCL25)	407000	407000 / 99

Granzyme B Mag Beads Lot#402913-000

Target Name	Bead Number	Concentration Std1	Standard
Granzyme B	51	21600 pg/ml	Standardmix H

Analytical information:

This product has been tested by Quality Control and passed internal specifications.

Quality control:

Jahins

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1–800–955–6288 (US and Canada) or 1–760–603–7200, x2 (all other countries). For inquiries, contact us at "thermofisher.com/askaquestion"