

Certificate of analysis

ProcartaPlex

 Cat. number: **EPX010-26010-901**
 96 tests/1 analyte

 Name: **ProcartaPlex Mouse I-TAC
(CXCL11) Simplex**

 Lot number: **436953-002**

 Expiry Date: **2026-08**

Components

S20837EX	Standardmix H
BD26085EX	I-TAC Mag Beads
BK26085EX	I-TAC Det. Ab.

Quantity	Lot	Store at
2 each	402510-000	2-8°C
1 x 120µl (50x)	406087-000	2-8°C
1 x 70µl (50x)	406088-000	2-8°C

Standardmix H Lot#402510-000

Provided below is a table of Standard 1 (Std1) value for each analyte in each tube when prepared according to the "Prepare antigen standard" procedure of the Manual. All data were evaluated in a 2-hour incubation at room temperature.

Analyte	Std1 Concentration (pg/ml)	ULOQ / LLOQ (pg/ml)
		Determined in Universal Assay Buffer
BLC (CXCL13)	8300	8300 / 2,03
CTACK (CCL27)	33700	8425 / 8,23
Eotaxin-2 (CCL24)	3550	3550 / 0,87
Granzyme B	23700	23700 / 5,79
IL-6R/sIL-6R	2462400	2462400 / 601
I-TAC (CXCL11)	10600	2650 / 2,59
TARC (CCL17)	3550	3550 / 0,87
TECK (CCL25)	514400	128600 / 126

I-TAC Mag Beads Lot#406087-000

Target Name	Bead Number	Concentration Std1	Standard
I-TAC (CXCL11)	57	10600 pg/ml	Standardmix H

Analytical information:

This product has been tested by Quality Control and passed internal specifications.

Quality control:

