

Certificate of Analysis



resDNASEQ® Quantitative Human DNA Kit

Product No. A26366
Lot No. 2505079
Date of Manufacture 28MAY2025
Expiration Date 26NOV2026

KIT CONFIGURATIONS:

PART NO.	COMPONENT DESCRIPTION	LOT NO.
100027320	BOX, resDNASEQ® Quantitative Human DNA Real-Time PCR Reagents	2505077
	4401975, (2) TUBE 2x Environmental Master Mix 2.0	2412140
	362250, TUBE Negative Control Water	2503189
	100027324, TUBE 10X Human DNA Real-Time PCR Assay Mix	2502070
100027323	BOX, resDNASEQ® Quantitative Human DNA Control	2505077
	100027325, TUBE Human DNA Control	2505070
	4405587, BTL DNA Dilution Buffer	2501123

TEST	SPECIFICATION	RESULT
Function Test	Positive result obtained when analysed on the Applied Biosystems 7500 FAST Real-Time PCR System using approximately 3 picograms of Human DNA Control per reaction.	Pass
Human DNA Control	Human DNA Control concentration is between 27.75 – 32.25 µg/mL	28.17
No Template Control	Less than 60 fg detected from No Template Control (NTC) reactions.	Pass

For Research Use Only. Not for use in diagnostic procedures.

Manufactured in compliance with our ISO 13485 certified quality management system. **Site:** Warrington, UK

Life Technologies
5781 Van Allen Way
Carlsbad, CA, USA 92008
www.thermofisher.com
For inquiries, contact us at
cofarequests@thermofisher.com


Quality Assurance
Issued 10JUN2025