

PRODUCT DATA SHEET

Pierce FlexMix Calibration Solution

Product Number: A39239

Lot Number: 3216582

Expiration: March 06, 2027

Summary:

The Thermo Scientific™ Pierce™ FlexMix Calibration Solution is a room-temperature stable, ready-to-use liquid formulation for the calibration of all Thermo Scientific Tribrid™ mass spectrometers. The only additional materials required are a 500µL Hamilton syringe and a liquid infusion line, both supplied with the instrument.

Visual:

Clear, colorless liquid

Components	Accurate Mass (m/z)	Specification (neat)	Results
*LC-MS grade acetonitrile (90% v/v)	--	Purity ≥ 99.9%	Pass
*LC-MS grade water (10% v/v)	--	LC-MS Grade	Pass
LC-MS grade acetic acid	59.0139 ^{b(-)}	Purity ≥ 99.7%	Pass
Imidazole	69.0447 ^{a(+)}	Purity ≥ 99.0%	Pass
Triethylamine	102.1277 ^{a(+)}	Purity ≥ 99.5%	Pass
Trifluoroacetic acid (TFA)	112.9856 ^{b(-)}	Purity ≥ 99.5%	Pass
Tetramethylpiperidine	142.1590 ^{a(+)}	Purity ≥ 97.5%	Pass
Pentafluoropropionic acid	162.9824 ^{b(-)}	Purity ≥ 97.0%	Pass
Caffeine	195.0876 ^{a(+)}	Purity ≥ 97.0%	Pass
Hexamethoxyphosphazene	322.0481 ^{a(+)}	Purity ≥ 99.0%	Pass
Perfluorooctanoic acid	362.9696 ^{b(-)}	Purity ≥ 96.0%	Pass
MRFA	524.2649 ^{a(+)}	Purity ≥ 98%	Pass
2,4,6-Tris(heptafluoropropyl)-1,3,5-triazine	601.9779 ^{c(-)}	Purity ≥ 95.0%	Pass
Hexakis(2,2-difluoroethoxy) phosphazene	622.0290 ^{a(+)}	Purity ≥ 97.0%	Pass
Hexakis(2,2,3,3-tetrafluoropropoxy) phosphazene	922.0098 ^{a(+)} , 1033.9881 ^{d(-)}	Purity ≥ 95.0%	Pass
Ultramark 1621	See note below	Purity ≥ 95.0%	Pass
Hexakis(1h,1h,7h-perfluorooheptoxy) phosphazene	2121.9331 ^{a(+)} , 2233.9115 ^{d(-)}	Purity ≥ 95.0%	Pass
Hexakis(1h,1h,9h-perfluorononyloxy) phosphazene	2721.8948 ^{a(+)} , 2833.8731 ^{d(-)}	Purity ≥ 95.0%	Pass

^a[M+H]⁺

^b[M-H]⁻

^c[M+OH]⁻

^d[M+TFA]⁻

*Combined concentration of all other components listed is less than 1% w/v

Note: Ultramark 1622 is a well-defined mixture of compounds producing a well-defined peak envelope in the range of 900-2200 m/z in the positive and negative ion modes.

Important Product Information:

- For best results, use a dedicated syringe and infusion line.
- Before and after calibration, flush the syringe and infusion line with 1:1 water/acetonitrile solution to clean.
- Turn on the mass spectrometer and wait for it to enter ready state.
- Infuse the calibration solution and optimize spray conditions until a spray stability of ≤ 15% RSD is achieved before initiating calibration routine.

Storage:

Upon receipt store at room temperature

Compatible with these Thermo Scientific™ Instruments:

Orbitrap Fusion™ MS

Orbitrap Fusion™ Lumos™ MS

Orbitrap ID-X MS

Orbitrap Tribrid™ Series 3.1 Instrument Control Software (or later)

Additional information:

For MSDS please visit

www.thermoscientific.com/pierce

For calibration procedure use:

Orbitrap Tribrid Series Getting Started Guide (June 2018)

Technical Support:

For hardware related support: 1-800-532-4752

For reagent support: 1-800-874-3723

Products are under warranty for one year from the date of shipment unless otherwise stated on this document or the product label.

Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED, INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON-INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to humans or animals...

Pierce Biotechnology
3747 N Meridian Road
Rockford, IL 61101

(800) 874-3723
(815) 968-0747

(815) 968-7316 fax
www.thermo.com/pierce

Analysis	Specification	Results
Positive Mode Target Ions (m/z) 69, 102, 142, 195, 322, 524, 622, 922, 1122, 1222, 1322, 1422, 1522, 1622, 1722, 1822, 2122, and 2722 m/z.	All calibration ions are present in the spectrum	Pass
Negative Mode Target Ions (m/z) 113, 163, 363, 602, 1034, 1234, 1334, 1434, 1534, 1634, 1734, 1834, 1934, 2234, and 2834 m/z	All calibration ions are present in the spectrum	Pass
Positive Mode Target Ions (m/z) 69, 102, 142, 195, 322, 524, 622, 922, 1522, 2122, and 2722 m/z	Ratio to Reference is between 0.5 and 2.0 for all ions listed below in positive ion mode: 69, 102, 142, 195, 322, 524, 622, 922, 1522, 2122, and 2722 m/z	Pass
Negative Mode Target Ions (m/z) 113, 163, 363, 602, 1034, 1534, 2234, and 2834 m/z	Ratio to Reference is between 0.5 and 2.0 for all ions listed below in negative ion mode: 113, 163, 363, 602, 1034, 1534, 2234, and 2834 m/z	Pass
Positive Mode Target Ions (m/z) 69, 102, 142, 195, 322, 524, 622, 922, 1122, 1222, 1322, 1422, 1522, 1622, 1722, 1822, 2122, 2722 Negative Mode Target Ions (m/z) 113, 163, 363, 602, 1034, 1234, 1334, 1434, 1534, 1634, 1734, 1834, 1934, 2234, 2834	Contaminant ions are not > 30% of the intensity of the target reference peak within +/- 10 m/z in both positive and negative mode*.	Pass

Publish Date: 22-May-2025
Timothy Morton, Sr. Quality Chemist, Mass Spectrometry



Products are under warranty for one year from the date of shipment unless otherwise stated on this document or the product label.

Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED, INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON-INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to humans or animals...


Pierce Biotechnology
3747 N Meridian Road
Rockford, IL 61101

(800) 874-3723
(815) 968-0747

(815) 968-7316 fax
www.thermo.com/pierce


Positive Ion Mode (Ion Trap, Normal Scan Mode), 50-220 m/z

07Mar2025_AB411434-D #1-5949 RT: 0.00-1.00 AV: 1298 NL: 1.53E6
F: ITMS + p ESI Full ms [50.0000-220.0000]



Positive Ion Mode (Ion Trap, Normal Scan Mode), 180-1000 m/z

07Mar2025_AB411434-D #1-5949 RT: 1.00-2.00 AV: 831 NL: 1.26E6
F: ITMS + p ESI Full ms [180.0000-1000.0000]



Products are under warranty for one year from the date of shipment unless otherwise stated on this document or the product label.

Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED, INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON-INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to humans or animals...


Pierce Biotechnology
3747 N Meridian Road
Rockford, IL 61101

(800) 874-3723
(815) 968-0747

(815) 968-7316 fax
www.thermo.com/pierce


Positive Ion Mode (Ion Trap, Normal Scan Mode), 800-2000 m/z

07Mar2025_AB411434-D #1-5949 RT: 2.00-3.00 AV: 666 NL: 1.67E6
F: FTMS + p ESI Full ms [800.0000-2000.0000]



Positive Ion Mode (Orbitrap, High Scan Mode), 1000-3000 m/z

07Mar2025_AB411434-D #1-5949 RT: 3.01-4.00 AV: 152 NL: 2.63E8
F: FTMS + p ESI Full ms [1000.0000-3000.0000]



Products are under warranty for one year from the date of shipment unless otherwise stated on this document or the product label.

Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED, INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON-INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to humans or animals...


Pierce Biotechnology
3747 N Meridian Road
Rockford, IL 61101

(800) 874-3723
(815) 968-0747

(815) 968-7316 fax
www.thermo.com/pierce


Negative Ion Mode (Ion Trap, Normal Scan Mode), 50-400 m/z

07Mar2025 AB411434-D #1-5949 RT: 4.50-5.50 AV: 1073 NL: 3.20E5
F: ITMS - p ESI Full ms [50.0000-400.0000]



Negative Ion Mode (Ion Trap, Normal Scan Mode), 280-1000 m/z

07Mar2025 AB411434-D #1-5949 RT: 5.50-6.50 AV: 847 NL: 3.28E5
F: ITMS - p ESI Full ms [280.0000-1000.0000]



Products are under warranty for one year from the date of shipment unless otherwise stated on this document or the product label.

Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED, INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON-INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to humans or animals...


Pierce Biotechnology
3747 N Meridian Road
Rockford, IL 61101

(800) 874-3723
(815) 968-0747


(815) 968-7316 fax
www.thermo.com/pierce

Negative Ion Mode (Ion Trap, Normal Scan Mode), 800-2000 m/z

07Mar2025_AB411434-D #1-5949 RT: 6.50-7.50 AV: 666 NL: 9.66E5
F: FTMS - p ESI Full ms [800.0000-2000.0000]



07Mar2025_AB411434-D #1-5949 RT: 7.51-8.50 AV: 152 NL: 8.80E8
F: FTMS - p ESI Full ms [1000.0000-3000.0000]



Negative Ion Mode (Orbitrap, High Scan Mode), 1000-3000 m/z

Products are under warranty for one year from the date of shipment unless otherwise stated on this document or the product label.

Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED, INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON-INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to humans or animals...

Pierce Biotechnology
3747 N Meridian Road
Rockford, IL 61101

(800) 874-3723
(815) 968-0747

(815) 968-7316 fax
www.thermo.com/pierce



Spectra were acquired on a Thermo Scientific™ Orbitrap Eclipse Mass Spectrometer

Products are under warranty for one year from the date of shipment unless otherwise stated on this document or the product label.

Products are warranted to operate or perform substantially in conformance with published Product specifications in effect at the time of sale, as set forth in the Product documentation, specifications and/or accompanying package inserts ("Documentation"). No claim of suitability for use in applications regulated by FDA is made. The warranty provided herein is valid only when used by properly trained individuals. Unless otherwise stated in the Documentation, this warranty is limited to one year from date of shipment when the Product is subjected to normal, proper and intended usage. This warranty does not extend to anyone other than Buyer. Any model or sample furnished to Buyer is merely illustrative of the general type and quality of goods and does not represent that any Product will conform to such model or sample. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE GRANTED, INCLUDING WITHOUT LIMITATION, IMPLIED WARRANTIES OF MERCHANTABILITY, FITNESS FOR ANY PARTICULAR PURPOSE, OR NON-INFRINGEMENT. BUYER'S EXCLUSIVE REMEDY FOR NON-CONFORMING PRODUCTS DURING THE WARRANTY PERIOD IS LIMITED TO REPAIR, REPLACEMENT OF OR REFUND FOR THE NON-CONFORMING PRODUCT(S) AT SELLER'S SOLE OPTION. THERE IS NO OBLIGATION TO REPAIR, REPLACE OR REFUND FOR PRODUCTS AS THE RESULT OF (I) ACCIDENT, DISASTER OR EVENT OF FORCE MAJEURE, (II) MISUSE, FAULT OR NEGLIGENCE OF OR BY BUYER, (III) USE OF THE PRODUCTS IN A MANNER FOR WHICH THEY WERE NOT DESIGNED, OR (IV) IMPROPER STORAGE AND HANDLING OF THE PRODUCTS. Unless otherwise expressly stated on the Product or in the documentation accompanying the Product, the Product is intended for research only and is not to be used for any other purpose, including without limitation, unauthorized commercial uses, in vitro diagnostic uses, ex vivo or in vivo therapeutic uses, or any type of consumption by or application to humans or animals...

Pierce Biotechnology
3747 N Meridian Road
Rockford, IL 61101

(800) 874-3723
(815) 968-0747

(815) 968-7316 fax
www.thermo.com/pierce