

# Certificate of Analysis



Applied Biosystems™ MYCOSEQ™ MYCOPLASMA PLUS Detection Kit

Product No. A55124  
Lot No. 2504015  
Date of Manufacture 22APR2025  
Expiration Date 06AUG2026

## KIT CONFIGURATIONS:

PART NO.	COMPONENT DESCRIPTION	LOT NO.
100112401	Box, MycoSeq™ Plus QPCR Reagent	2504016
	362250 – Negative Control Water	2411188
	100112365 - Tube, MycoSeq™ Plus QPCR 10X Assay Mix	2503011
	100112399 - Tube, 2X Master Mix Plus	3201777
100112402	Box, MycoEQ™ Plus QPCR DNA Control	2504014
	4405587 - Btl, DNA Dilution Buffer	2501123
	100113960 - Tube, MycoSeq™ Plus Discriminatory Positive Control	2503011

TEST	SPECIFICATION	RESULT
------	---------------	--------

QC Method: The positive results obtained when analysed on the Fast 7500 Real Time PCR system using approximately 2000 copies of DNA control ( P/N 100113960) per reaction are the following:

FAM (DPC) Average Ct Pass/Fail PASS

VIC (DPC) Average Ct Pass/Fail PASS

NED (IPC) Average Ct Pass/Fail PASS

QC Method: The negative results obtained when analysed on the Fast 7500 Real Time PCR system without DNA:

No-Template Control (NTC) Pass/Fail PASS

*For Research Use Only. Not for use in diagnostic procedures.*

Manufactured in compliance with our ISO 13485 certified quality management system.

Site: Warrington

Life Technologies

5781 Van Allen Way

Carlsbad, CA, USA 92008

[www.thermofisher.com](http://www.thermofisher.com)

For inquiries, contact us at

[cofarequests@thermofisher.com](mailto:cofarequests@thermofisher.com)

Quality Assurance  
Issued 30APR2025