

Certificate of analysis

ProcartaPlex

 Cat. number: **EPX010-12282-901**

96 tests/1 analyte

 Name: **ProcartaPlex Human C3a Simplex**

 Lot number: **446240-002**

 Expiry Date: **2028-06**

Components

Components	Quantity	Lot	Store at
BD12282EX C3a Mag Beads	1 x 120µl (50x)	441871-000	2-8°C
BK12282EX C3a Det. Ab.	1 x 70µl (50x)	441872-000	2-8°C
S12282EX C3a Standard Mix	2 each	408962-000	2-8°C
UAB10x/11 Universal Assay Buffer 10x	1 x 10ml	435317-000	2-8°C

C3a Mag Beads Lot#441871-000

Target Name	Bead Number	Concentration Std1	Standard
C3a	14	101600 pg/ml	C3a Standard Mix

C3a Standard Mix Lot#408962-000

Provided below is a table of Standard 1 (Std1) value for each analyte in each tube when prepared according to the "Prepare antigen standard" procedure of the Manual. All data were evaluated in a 2-hour incubation at room temperature.

Analyte	Std1 Concentration (pg/ml)	ULOQ / LLOQ (pg/ml)
C3a	101600	Determined in cell culture medium 25400 / 25

Analytical information:

This product has been tested by Quality Control and passed internal specifications.

Quality control:

For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries). For inquiries, contact us at "thermofisher.com/askaquestion"