

Certificate of analysis

ProcartaPlex

Cat. number:

11 analytes

Name:

Canine StandardLot number: **454490-000**Expiry Date: **2026-06**

Components

S50512EX 11-plex Standard

Quantity

2 each

Lot

333518-000

Store at

2-8°C

11-plex Standard Lot#333518-000

Provided below is a table of Standard 1 (Std1) value for each analyte in each tube when prepared according to the "Prepare antigen standard" procedure of the Manual. All data were evaluated in a 2-hour incubation at room temperature.

Analyte	Std1 Concentration (pg/ml)	ULOQ / LLOQ (pg/ml)
		Determined in cell culture medium
IFN gamma	15300	15300 / 3,74
IL-10	21900	5475 / 5,35
IL-12/IL-23p40	23200	23200 / 5,66
IL-2	29300	29300 / 7,15
IL-6	52700	52700 / 13
IL-8 (CXCL8)	33200	33200 / 8,11
MCP-1 (CCL2)	20900	20900 / 5,10
NGF beta	16700	16700 / 4,08
SCF	20800	20800 / 5,08
TNF alpha	16400	16400 / 4,00
VEGF-A	14800	14800 / 3,61

Analytical information:

This product has been tested by Quality Control and passed internal specifications.

Quality control:



For Research Use Only. Not for use in diagnostic procedures. If you have any further questions about this Certificate of Analysis, please contact Technical Services at 1-800-955-6288 (US and Canada) or 1-760-603-7200, x2 (all other countries).
For inquiries, contact us at "thermofisher.com/askaquestion"