

## Quality Assurance Certificate

**Minisart® NML**

**Order no.** 17597-----Q  
**Filter material** SFCA  
**Use before** 08 / 2029  
**Pore size** 0.20 µm  
**Lot no.** 240975103

Minisart Syringe Filters are manufactured under a Quality Management System certified by an accredited Notified Body according to the EN ISO 13485 Standard for Medical Devices and to ISO 9001 Quality System Standard.



Each unit is tested 100% during manufacture by camera check and housing integrity by a leakage test under automatic conditions. Before packing, samples of each lot were checked by visual inspection.

Minisart is biocompatible according to ISO 10993 series.

In addition, Minisart is CE marked according to Medical Device Directive 93/42/EEC Annex II, class IIa.

Each lot has been sampled, tested and released by Quality Department for the following specifications:

*Burst Pressure Test (In-Process Control)*

*Bubble Point Test*

*Pressure Hold Test (In-Process Control)*

*Flow Rate Performance*

*Sterile Filtration Capability (Bacterial Challenge Test)*

Dr. Anna Vreemann  
Site Quality Manager

Manufactured by Sartorius Stedim Biotech GmbH  
37070 Goettingen, Germany

Phone +49.551.308.0  
Fax +49.551.308.3289

[www.sartorius.com](http://www.sartorius.com)