

## Quality Assurance Certificate

Vivaspin® 500

**Order no.** VS0191  
**Use before** 04 / 2027  
**Pore size** 3K MWCO PES  
**Lot no.** 251900797

This document is to verify that the designated product was manufactured by Sartorius Stedim according to a Quality Management System that is certified for compliance with current ISO 9001.

Sartorius hereby declares that the designated product was manufactured to meet the published specifications for these products.

Sartorius also declares that the designated products have been manufactured from approved materials of construction and comply with design and manufacturing documentation.

The department of quality control verifies that the designated products have each been tested for integrity using the pressure hold test method, and a representative sample have been tested for retention using a standard molecular weight marker solution.

The product has passed Sartorius Stedim's in house tests and thus meets Sartorius Stedim's stringent quality control standards.

29. May 2025

Date



Anthony Packham  
Site Quality Manager

Manufactured by Sartorius Stedim Lab Ltd.  
GL 10 3UT Stonehouse, Gloucestershire, UK

Phone +44.1453.821972  
Fax +44.1453.827928

[www.sartorius.com](http://www.sartorius.com)