

Quality Assurance Certificate

Minisart® NML

Order no. 16592-----K
Use before 02 / 2028
Pore size 0.80 µm
Lot no. 250345103

Minisart is manufactured by Sartorius in accordance with established quality standards and in a facility whose Quality Management System is certified for compliance with the DIN EN ISO 9001 Quality System Standard and DIN EN ISO 13485.

This product was sterilized using a validated process following DIN/EN ISO 11135 regulations.

Each unit is tested 100% during manufacture for membrane and housing integrity by a leakage test under automatic conditions. Before packing each unit has been checked by visual inspection.

Minisart is biosafe according to the USP Class VI Biological tests for Plastics. It also complies with the Title 21 Code of Federal Regulations, Section 210.3(b)(6) and 211.72 for non-fiber-releasing filters.

Following 100% control, each lot has been sampled, tested and released by QA Department for the following characteristics:

Burst Pressure Test

Bubble Point Test

Endotoxine Test

Sterility

Pressure Hold Test

Flow Rate Performance

2025-02-13

Date



Dr. Anna Vreemann
Site Quality Manager

Manufactured by Sartorius Stedim Biotech GmbH
37070 Goettingen, Germany

Phone +49.551.308.0
Fax +49.551.308.3289

www.sartorius.com