

Quality Assurance Certificate

Vivacon® 500
PCR Grade

Order no. VN01H42ETO
Use before 08 / 2026
Lot no. 233700757

This document is to verify that the designated product was manufactured by Sartorius according to a Quality Management System that is certified for compliance with BS EN ISO 9001:2015.

This product has been subjected to and has fulfilled Sartorius Stedim Biotech's quality control standards from raw material to the final product.

Each manufacturing lot is sampled, tested and released by Quality Assurance with respect to the following characteristics:

Physical inspection:
Products are 100% inspected in accordance with their current applicable specification.

Performance test:
Product samples are assayed for flow rate and membrane retention characteristics in accordance with their current applicable specification.

Notes:**Sterility:**

The products have been sterilised by dual-cycle ethylene oxide gas.

DNA Free:

The products are certified to be free from detectable human genomic DNA as assayed by quantitative PCR amplification where the limit of detection is ≤ 2.0 pg.

31. January 2024

Date



Anthony Packham
Site Quality Manager

Manufactured by Sartorius Stedim Lab Ltd.
GL 10 3UT Stonehouse, Gloucestershire, UK

Phone +44.1453.821972
Fax +44.1453.827928

www.sartorius.com