

Certificate of Analysis

Product Name: MCFARLAND STD SET- 0.5-1-2-3-4
Product Code: R20421
Lot Number: 6249825
Expiration Date: 13.07.2026
Date of Manufacture: 22.05.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W77BCL06G	McFarland Standard 0.5 Filled Bulk	6249831	13.07.2026
W77BCL07H	McFarland Standard 1.0 Filled Bulk	6249833	14.07.2026
W77BCL08J	McFarland Standard 2.0 Filled Bulk	6249835	14.07.2026
W77BCL09K	McFarland Standard 3.0 Filled Bulk	6249837	14.07.2026
W77BCL10L	McFarland Standard 4.0 Filled Bulk	6249839	14.07.2026

McFarland Equivalence Turbidity Standards are designed to be visual standards for determining approximate microorganism concentration.

A 0.5 McFarland Standard is approximately equivalent to 150 million cfus/ml.

A 1.0 McFarland Standard is approximately equivalent to 300 million cfus/ml.

A 2.0 McFarland Standard is approximately equivalent to 600 million cfus/ml.

A 3.0 McFarland Standard is approximately equivalent to 900 million cfus/ml.

A 4.0 McFarland Standard is approximately equivalent to 1.2 billion cfus/ml.

Variation in organisms' size and phase of log growth will affect organism density.

*McFarland Equivalence 0.5 is defined by CLSI as 0.08 to 0.10 absorbancy at 625 nm.



Fay Curd
Quality Assurance Manager