

## **Certificate of Analysis**

Product Name: MCFARLAND STD SET- 0.5-1-2-3-4

Product Code: R20421 Lot Number: 6268782 Expiration Date: 21.09.2026 Date of Manufacture: 30.07.2025

\*Date format is DD.MM.YYYY

| ltem      | Description                        | Lot/Serial | Expire     |
|-----------|------------------------------------|------------|------------|
| W77BCL06G | McFarland Standard 0.5 Filled Bulk | 6292897    | 21.09.2026 |
| W77BCL07H | McFarland Standard 1.0 Filled Bulk | 6261137    | 21.09.2026 |
| W77BCL08J | McFarland Standard 2.0 Filled Bulk | 6261141    | 21.09.2026 |
| W77BCL09K | McFarland Standard 3.0 Filled Bulk | 6261145    | 21.09.2026 |
| W77BCL10L | McFarland Standard 4.0 Filled Bulk | 6261149    | 21.09.2026 |

McFarland Equivalence Turbidity Standards are designed to be visual standards for determining approximate microorganism concentration.

A 0.5 McFarland Standard is approximately equivalent to 150 million cfus/ml.

A 1.0 McFarland Standard is approximately equivalent to 300 million cfus/ml.

A 2.0 McFarland Standard is approximately equivalent to 600 million cfus/ml.

A 3.0 McFarland Standard is approximately equivalent to 900 million cfus/ml.

A 4.0 McFarland Standard is approximately equivalent to 1.2 billion cfus/ml.

Variation in organisms' size and phase of log growth will affect organism density.

\*McFarland Equivalence 0.5 is defined by CLSI as 0.08 to 0.10 absorbancy at 625 nm.

Spo-

Fay Curd Quality Assurance Manager