

Certificate of Analysis

Product Name: McFarland Turb Standard 1.0
Product Code: R20411
Lot Number: 6261127
Expiration Date: 04.08.2026
Date of Manufacture: 01.07.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W77BCL07H	McFarland Standard 1.0 Filled Bulk	6261136	04.08.2026

McFarland Equivalence Turbidity Standards are designed to be visual standards for determining approximate microorganism concentration.

A 1.0 McFarland Standard is approximately equivalent to 300 million cfus/ml. Variation in organisms size and phase of log growth will affect organism density.



Fay Curd
Quality Assurance Manager