

CERTIFICATE OF ANALYSIS

PRODUCT CM0131B
TRYPTONE SOYA AGAR 500g

LOT NUMBER 6279163

EXPIRY DATE 2030.06.30

DATE OF MANUFACTURE 2025.06.09

Delivery/Customer information

Date Printed

2025.07.01

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Appearance	Straw powder	Straw powder
Colour on reconstitution	Straw 2	Straw 1-2
pH (25°C)	7.3	7.1 - 7.5
Clarity	Clear	Clear
Blood compatibility with horse blood	Satisfactory	Satisfactory
Blood compatibility with sheep blood	Satisfactory	Satisfactory

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

Aerobic incubation at 30-35°C for 18-24 hours

<i>Streptococcus pyogenes</i> ATCC®19615	25	32	128	Pale straw cols
<i>Streptococcus viridans</i> NCTC1080	25	24	96	Pale straw cols
<i>Staphylococcus aureus</i> ATCC®9144	58	55	95	Straw cols
<i>Staphylococcus epidermidis</i> ATCC®12228	61	49	80	White/grey cols

Enriched with 7% v/v horse blood

Aerobic incubation at 37°C for 24 hours

<i>Streptococcus pyogenes</i> ATCC®19615	52	63	121	Pale straw cols, beta haemolysis
<i>Streptococcus viridans</i> NCTC1080	45	46	102	Grey/green cols, alpha haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	41	52	127	Grey/green cols, alpha haemolysis

Enriched with 7% v/v horse blood

Microaerophilic incubation at 37°C for 48 hours

<i>Haemophilus influenzae</i> ATCC®19418	35	40	114	Colourless cols
--	----	----	-----	-----------------

Growth of *Haemophilus* spp. in the presence of X & V factor discs

Aerobic incubation at 37°C for 18 hours

	X Factor (mm)	V Factor (mm)	X+V Factor (mm)
<i>Haemophilus influenzae</i> ATCC®9334	0	0	17
<i>Haemophilus influenzae</i> ATCC®19418	0	0	21
<i>Haemophilus influenzae</i> ATCC®49247	0	0	17
<i>Haemophilus parainfluenzae</i> ATCC®33392	0	41	48

Enriched with 7% v/v horse blood

Bacitracin activity

Aerobic incubation at 37°C for 18 hours

CERTIFICATE OF ANALYSIS

PRODUCT CM0131B
TRYPTONE SOYA AGAR 500g
LOT NUMBER 6279163

	Zone Size (mm)	Limits (mm)
<i>Streptococcus pyogenes</i> ATCC®19615	16	10 - 20

Tested in accordance with current USP/EP/BP/JP

Aerobic incubation at 30-35°C for 24 hours

<i>Staphylococcus aureus</i> ATCC®6538	33	40	121	Straw cols
<i>Escherichia coli</i> ATCC®8739	45	52	116	Cream cols
<i>Bacillus subtilis</i> ATCC®6633	62	53	85	Irregular, straw cols
<i>Pseudomonas aeruginosa</i> ATCC®9027	65	65	100	Straw cols
<i>Salmonella typhimurium</i> ATCC®14028	55	48	87	Straw cols
<i>Salmonella abony</i> NCTC6017	91	79	87	Straw cols

Aerobic incubation at 30-35°C for up to 5 days

<i>Candida albicans</i> ATCC®10231	55	58	105	Cream cols
<i>Aspergillus brasiliensis</i> ATCC®16404	41	48	117	White mycelia, no spores

Testing has been performed in accordance with ISO11133:2014

Aerobic incubation at 30 ± 2°C for 24 ± 2 hours

<i>B. cereus</i> ATCC®11778 WDCM00001	109	128	118	Irregular, straw cols
<i>B. subtilis</i> ATCC®6633 WDCM00003	62	53	85	Irregular, straw cols
<i>E. coli</i> ATCC®8739 WDCM00012	89	103	116	Cream cols

Aerobic incubation at 36 ± 2°C for 20 ± 2 hours

<i>E. coli</i> ATCC®25922 WDCM00013	53	51	96	Cream cols
<i>E. coli</i> ATCC®11175 WDCM00090	58	53	91	Cream cols
<i>E. coli</i> NCTC13167 WDCM00179	53	45	85	Cream cols
<i>P. aeruginosa</i> ATCC®10145 WDCM00024	106	129	122	Straw cols
<i>E. faecalis</i> ATCC®29212 WDCM00087	60	75	125	Straw cols

Aerobic incubation at 37 ± 2°C for 24 ± 2 hours

<i>S. aureus</i> ATCC®25923 WDCM00034	71	61	85	Straw cols
<i>L. monocytogenes</i> ATCC®13932 WDCM00021	63	61	98	Straw cols

Aerobic incubation at 44 ± 2°C for 21 ± 3 hours

<i>E. coli</i> ATCC®8739 WDCM00012	50	51	101	Cream cols
------------------------------------	----	----	-----	------------

Anaerobic incubation at 44 ± 2°C for 21 ± 3 hours

<i>C. perfringens</i> ATCC®13124 WDCM00007	66	73	111	Straw cols
--	----	----	-----	------------

CERTIFICATE OF ANALYSIS

PRODUCT CM0131B
TRYPTONE SOYA AGAR 500g
LOT NUMBER 6279163

Tested in accordance with current CLSI M22 A

Enriched with 5% v/v sheep blood

Aerobic incubation at 35°C for 18-24 hours

<i>Streptococcus pyogenes</i> ATCC®19615	1E+04-1E+06	Pale straw cols, beta haemolysis
<i>Streptococcus pneumoniae</i> ATCC®6305	1E+04-1E+06	Grey/green cols, alpha haemolysis
<i>Staphylococcus aureus</i> ATCC®25923	1E+04-1E+06	White/grey cols
<i>Escherichia coli</i> ATCC®25922	1E+04-1E+06	Straw cols

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.
Refer to product specification for full details.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.07.01

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.