

CERTIFICATE OF ANALYSIS

PRODUCT DD0013T
ONPG SINGLE DISCS

LOT NUMBER 6245712

EXPIRY DATE 2026.03.19

DATE OF MANUFACTURE 2025.03.19

Delivery/Customer information

Date Printed

2025.05.23

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Initial Inoculum

Test Result

Diagnostic characteristics

Escherichia coli ATCC®25922

1E+04-1E+06

Yellow

Citrobacter freundii ATCC®8090

1E+04-1E+06

Yellow

Salmonella enteritidis ATCC®13076

1E+04-1E+06

Colourless

Shigella flexneri ATCC®12022

1E+04-1E+06

Colourless

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.05.23

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com