

Certificate of Quality

**Product Name:** KBE (Kanamycin Bile Esculin) Agar 10/PK  
**Lot Number:** 281186


**Product Number:** R01510  
**Expiration Date:** 2025-08-18  
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

| Organism Challenge           | ATCC  | Performance Criteria    |
|------------------------------|-------|-------------------------|
| Bacteroides thetaiotaomicron | 29148 | Good Growth, Blackening |
| Bacteroides fragilis         | 25285 | Good Growth, Blackening |
| Clostridium perfringens      | 13124 | No Growth               |
| Escherichia coli             | 25922 | No Growth               |
| Staphylococcus aureus        | 25923 | No Growth               |

**Physical Characteristics:**  
Appearance: Beige-Green, Clear Agar  
pH @ 20-25°C: N/A

**Microbial Load Evaluation:**  
Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed  
  
Senior Quality Assurance Engineer