

Certificate of Quality

Product Name: ALAMAR BLUE 100ML
Lot Number: 255315

Product Number: 00-100
Expiration Date: 2026-03-31
(YYYY-MM-DD)

This product has been manufactured, processed and packaged in accordance with Quality Systems Regulation, 21 CFR Part 820. Representative samples were tested per Remel Inc., a part of Thermo Fisher Scientific Quality Control specifications and were found to meet performance criteria for this product.

Absorbance Range and Correlation:

Genesys 10S VIS	0.84-0.95 AU
Perkin Elmer Model Lambda 2	0.87-0.97 AU


Physical Characteristics:

Appearance: Dark Blue Solution With No Visible Precipitate

Microbial Load Evaluation:

Manufactured lot has been statistically sampled. Samples were acceptable for microbial content and aesthetic value following a minimum of 72 hours at 33-37°C.

Signed



Senior Quality Assurance Engineer