

CERTIFICATE OF ANALYSIS

PRODUCT	PB5012A	TRYPTONE SOYA AGAR + 5% SHEEP BLOOD
LOT NUMBER	6278524	
EXPIRY DATE	2025.08.07	
TEST DATE	2025.06.12	
REPORTING DATE	2025.06.16	

General Characteristics	Results	Specification
Colour	Conforms	Traffic red
Appearance	Conforms	Opaque
pH	7.5	7.3 -7.7
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®6538	65	59	2-4mm, yellow shiny cols with haemolysis
Bacillus subtilis ATCC®6633	57	42	4-8 mm, light grey, rough irregular cols
Bacillus cereus ATCC®11778	63	51	5-8mm, rough cols with double hemolysis
Pseudomonas aeruginosa ATCC®9027	102	114	3-8 mm, grey shiny colonies
Streptococcus pyogenes ATCC®12344	<10000	Conforms	Good growth, grey cols & beta haemolysis
Streptococcus pneumoniae ATCC®6305	<10000	Conforms	Good growth, grey cols & alpha haemolysis

For positive strains, colony count is greater than or equal to 50% of the control medium.

Tested in accordance with BP/EP/JP/USP. For positive strains of B. cereus colony count must be 50-200% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.