

## CERTIFICATE OF ANALYSIS

**PRODUCT** SR0054C  
EGG YOLK TELLURITE EMULSION

**LOT NUMBER** 6254207

**EXPIRY DATE** 2026.04.30

**DATE OF  
MANUFACTURE** 2025.04.07

## Delivery/Customer information

Date Printed

2025.04.25

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Colour	Yellow	Yellow
Clarity	Opaque	Opaque

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

## Tested in Baird-Parker Agar Base CM0275

Aerobic incubation at  $37 \pm 2^{\circ}\text{C}$  for  $48 \pm 2$  hours

<i>Staphylococcus aureus</i> ATCC®9144	52	74	141	Shiny black cols, white & clear zones
--	----	----	-----	---------------------------------------

## Inoculation using diminishing sweep technique

<i>Staphylococcus epidermidis</i> ATCC®14990	>1E+04	-	-	Black cols, no zones
<i>Proteus mirabilis</i> ATCC®29906	>1E+04	-	-	Brown/black cols, no zones

## Testing has been performed in accordance with ISO11133:2014

Aerobic incubation at  $37 \pm 2^{\circ}\text{C}$  for  $48 \pm 2$  hours

<i>S. aureus</i> ATCC®25923 WDCM00034	108	141	130	Shiny black cols, white & clear zones
<i>S. aureus</i> ATCC®6538 WDCM00032	115	130	114	Shiny black cols, white & clear zones

## Inoculation using diminishing sweep technique

<i>S. saprophyticus</i> ATCC®15305 WDCM00159	3.5E+03	-	-	Black cols, no zones
<i>S. epidermidis</i> ATCC®12228 WDCM00036	3.8E+03	-	-	Black cols, weak zone
<i>E. coli</i> ATCC®25922 WDCM00013	>1E+04	0	-	No growth
<i>E. coli</i> ATCC®8739 WDCM00012	>1E+04	0	-	No growth

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.  
Refer to product specification for full details.

**CERTIFICATE OF ANALYSIS**

<b>PRODUCT</b>	<b>SR0054C</b>
	<b>EGG YOLK TELLURITE EMULSION</b>
<b>LOT NUMBER</b>	<b>6254207</b>

---

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.04.25

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

**OXOID LIMITED**

**Wade Road, Basingstoke, Hampshire RG24 8PW. England**

[www.ThermoFisher.com](http://www.ThermoFisher.com)