

CERTIFICATE OF ANALYSIS

PRODUCT SR0054C
EGG YOLK TELLURITE EMULSION

LOT NUMBER 6287813

EXPIRY DATE 2026.07.31

**DATE OF
MANUFACTURE** 2025.07.01

Delivery/Customer information

Date Printed

2025.07.21

Delivery No.

Customer

Customer Order Number

Physical Characteristics	Results	Specification
Colour	Yellow	Yellow
Clarity	Opaque	Opaque

Microbiological Performance	Control cfu	Test cfu	Recovery Test %	Description
-----------------------------	----------------	-------------	--------------------	-------------

Tested in Baird-Parker Agar Base CM0275

Aerobic incubation at $37 \pm 2^{\circ}\text{C}$ for 48 ± 2 hours

<i>Staphylococcus aureus</i> ATCC®9144	69	63	91	Shiny black cols,white & clear zones
--	----	----	----	--------------------------------------

Inoculation using diminishing sweep technique

<i>Staphylococcus epidermidis</i> ATCC®14990	>1E+04	-	-	Black cols, no zones
<i>Proteus mirabilis</i> ATCC®29906	>1E+04	-	-	Brown/black cols, no zones

Testing has been performed in accordance with ISO11133:2014

Aerobic incubation at $37 \pm 2^{\circ}\text{C}$ for 48 ± 2 hours

<i>S. aureus</i> ATCC®25923 WDCM00034	107	106	99	Shiny black cols,white & clear zones
<i>S. aureus</i> ATCC®6538 WDCM00032	108	100	93	Shiny black cols,white & clear zones

Inoculation using diminishing sweep technique

<i>S. saprophyticus</i> ATCC®15305 WDCM00159	2.8E+03	-	-	No growth
<i>S. epidermidis</i> ATCC®12228 WDCM00036	2.9E+03	-	-	Black cols, no zones
<i>E. coli</i> ATCC®25922 WDCM00013	>1E+04	0	-	No growth
<i>E. coli</i> ATCC®8739 WDCM00012	>1E+04	0	-	No growth

Control Medium: Tryptone Soya Agar

A satisfactory result is represented by recovery of positive strains equal to or greater than 70% of the control medium.
Refer to product specification for full details.

CERTIFICATE OF ANALYSIS

PRODUCT	SR0054C
	EGG YOLK TELLURITE EMULSION
LOT NUMBER	6287813

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.07.21

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com