

## CERTIFICATE OF ANALYSIS

PRODUCT	PO5163A	M-CP AGAR (MEMB.CLOSTRIDIUM PERFRINGENS)
LOT NUMBER	6270878	
EXPIRY DATE	2025.07.09	
TEST DATE	2025.05.14	
REPORTING DATE	2025.05.19	

General Characteristics	Results	Specification
Colour	Conforms	Blue lilac
Appearance	Conforms	Transparent
pH	7.8	7.4 -7.8
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Clostridium perfringens ATCC®13124 WDCM 00007	<10000	Conforms	Good growth, yellow cols, Phosphat. +
Clostridium bifermentans NCTC506 WDCM 00079	<10000	Conforms	Good growth, blue colonies, Phosphat. -
Escherichia coli ATCC®25922 WDCM 00013	>100,000	Conforms	Total inhibition

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.