

CERTIFICATE OF ANALYSIS

PRODUCT	PB5008A	COLUMBIA AGAR + SHEEP BLOOD
LOT NUMBER	6291712	
EXPIRY DATE	2025.09.01	
TEST DATE	2025.07.07	
REPORTING DATE	2025.07.10	

General Characteristics	Results	Specification
Colour	Conforms	Carmine red
Appearance	Conforms	Opaque
pH	7.5	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Streptococcus pyogenes ATCC®12344	99	95	1mm, light-grey cols, beta-haemolysis
Staphylococcus aureus ATCC®25923	<10000	Conforms	Good growth, white colonies
Staphylococcus aureus ATCC®6538	<10000	Conforms	Good growth, yellow cols & haemolysis
Streptococcus pneumoniae ATCC®6305	<10000	Conforms	Good growth, grey cols & alpha haemolysis
Escherichia coli ATCC®8739	<10000	Conforms	Good growth, dark grey colonies

For positive strains, colony count is greater than or equal to 70% of the control medium.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.