

## CERTIFICATE OF ANALYSIS

Page 1 of 1

PRODUCT	PO5138A HAEMOPHILUS TEST MEDIUM (HTM)
LOT NUMBER	6278603
EXPIRY DATE	2025.08.19
TEST DATE	2025.06.13
REPORTING DATE	2025.06.16

General Characteristics	Results	Specification
Colour	Conforms	Sand yellow
Appearance	Conforms	Transparent
pH	7.5	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Susceptibility Testing	Results(mm)	Limits (mm)
Haemophilus influenzae ATCC®49247 - MXF5	31	31 -39
Haemophilus influenzae ATCC®49247 -CLR15	16	11 -17
Haemophilus influenzae ATCC®49247 -CPD10	31	25 -31

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133:2014.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133:2014 and registered under D-PL-20190-01-00.