

Certificate of Analysis

Product Name: DRYSPOT PNEUMO 60 TESTS
Product Code: DR0420M
Lot Number: 6235448
Expiration Date: 28.02.2027
Date of Manufacture: 20.05.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74DR424D	Strep Pneu Test Rgt Dryspot	6184789	28.02.2027
W74DR427G	Dryspot Pneu Neg Con	6184790	28.02.2027
W74DR429A	Pneumo Diluent Buffer	6184791	28.02.2027

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

Organism Challenge	OCC	ATCC	NCTC	Performance Criteria
S. pneumoniae	687	6303		Acceptable
S. pneumoniae	1069	11887		Acceptable
S. pneumoniae	1107	11899		Acceptable
S. pneumoniae	1109	11901		Acceptable
S. pneumoniae	1165	11908		Acceptable
A. viridans	1660	11563	8251	Acceptable
Streptococcus B1	761	9993		Acceptable
Streptococcus F5	971			Acceptable
Streptococcus G5	987			Acceptable
Streptococcus mutans	690	25175	10449	Acceptable

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.



Fay Curd
Quality Assurance Manager