

## Certificate of Analysis

**Product Name:** Staphaurex Plus KT/150TST  
**Product Code:** R30950102  
**Lot Number:** 6186662  
**Expiration Date:** 04.04.2026  
**Date of Manufacture:** 27.05.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74CZL33E	Staph Plus Control Latex Filled Bulk	6196412	04.04.2026
W75DZL33F	Staph Plus Test Latex Filled Bulk	6196413	04.04.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Smoothness	Test and Control latex must not show agglutination although a trace granularity is acceptable	Conforms
	Test and Control Latex should spread easily to cover the circle	Conforms
Efficacy/ Specificity	Test Latex must show obvious agglutination with the appropriate positive Staphylococcus cultures	Conforms
	Control Latex must not show agglutination	Conforms
	Test and Control Latex must not show agglutination with the appropriate negative Staphylococcus cultures (trace granularity is acceptable)	Conforms



Fay Curd  
Quality Assurance Manager