

## Certificate of Analysis

**Product Name:** Streptex KT/50TST  
**Product Code:** R30950501  
**Lot Number:** 6227221  
**Expiration Date:** 31.05.2026  
**Date of Manufacture:** 10.06.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W73UZL55T	ZL55 Extraction Enzyme Freeze dried	6192757	17.02.2028
W75UZL56N	ZL56 Strep F Latex Filled Bulk	6190211	10.06.2026
W75UZL54S	ZL54 Strep D Latex Filled Bulk	6217497	12.08.2026
W74UZL58G	Streptex Pos. Control Filled Bulk	6192760	03.03.2027
W75AZL51A	ZL51 Strep A Latex	6222057	12.08.2026
W75BZL52W	ZL52 Strep B Latex Filled Bulk	6190214	31.05.2026
W75WZL53W	ZL53 Strep C Latex Filled Bulk	6197687	09.07.2026
W75WZL57L	ZL57 Strep G Latex Filled Bulk	6193181	05.06.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Abbreviated Latex Efficacy/Specificity	Each Latex must show obvious agglutination with its homologous extract in the 10 and 60 minute procedures	Conforms
Reactivity of Latex with Polyvalent Positive Control & Extraction Enzyme	The Polyvalent Positive Control must show agglutination of at least ++ with each latex and the reactions must be of roughly equivalent strength	Conforms
	The Extraction Enzyme must show no agglutination with the Latexes, although a trace reaction is acceptable	Conforms



Fay Curd  
Quality Assurance Manager