

Certificate of Analysis

Product Name: Streptex KT/50TST
Product Code: R30950501
Lot Number: 6211471
Expiration Date: 21.04.2026
Date of Manufacture: 07.05.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W73UZL55T	ZL55 Extraction Enzyme Freeze dried	6192757	17.02.2028
W75UZL56N	ZL56 Strep F Latex Filled Bulk	6190211	10.06.2026
W74UZL58G	Streptex Pos. Control Filled Bulk	6192760	03.03.2027
W75WZL53W	ZL53 Strep C Latex Filled Bulk	6197687	09.07.2026
W75AZL51A	ZL51 Strep A Latex	6190207	21.04.2026
W75BZL52W	ZL52 Strep B Latex Filled Bulk	6190214	31.05.2026
W75UZL54S	ZL54 Strep D Latex Filled Bulk	6190213	07.05.2026
W75WZL57L	ZL57 Strep G Latex Filled Bulk	6193181	05.06.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Abbreviated Latex Efficacy/Specificity	Each Latex must show obvious agglutination with its homologous extract in the 10 and 60 minute procedures	Conforms
Reactivity of Latex with Polyvalent Positive Control & Extraction Enzyme	The Polyvalent Positive Control must show agglutination of at least ++ with each latex and the reactions must be of roughly equivalent strength	Conforms
	The Extraction Enzyme must show no agglutination with the Latexes, although a trace reaction is acceptable	Conforms



Fay Curd
Quality Assurance Manager