

Certificate of Analysis

Product Name: LEGIONELLA SPECIES TEST KIT 1 X 50 TESTS
Product Code: DR0803M
Lot Number: 6272867
Expiration Date: 26.09.2026
Date of Manufacture: 05.09.2025

*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W74DR803M	legionella Other species Latex Reagent	6272869	26.09.2026
W74DR806M	Legionella Control Latex	6272872	26.09.2026
W74DR807M	Legionella Dil Buffer	6272873	28.09.2026

The above have been tested in our Quality Control Laboratory using conventional procedures and control ATCC and NCTC cultures where appropriate and have met the specified test parameters. Additional challenging strains may also be employed.

Organism Challenge	OCC	ATCC	NCTC	Performance Criteria
L pneumophila	1273	43111	11404	Acceptable
L pneumophila	1282	43113	12098	Acceptable
L pneumophila	1157	12821		Acceptable
L pneumophila	1274	43130		Acceptable
L pneumophila	1275	43290		Acceptable
L pneumophila	1276	43736		Acceptable
L pneumophila	1197	33156	11233	Acceptable
L pneumophila	1196	33155	11232	Acceptable
L pneumophila	1286	35251		Acceptable
L pneumophila	789	33215	11406	Acceptable
L micadadei	1280	33204		Acceptable
L anisa	1289		11974	Acceptable
L dumofii	1294	35850		Acceptable
L gormanii	1296	43769		Acceptable
L rubrilucens	1215	35304	11987	Acceptable
L jamestowniensis	1230	35298	11981	Acceptable
L moravica	1281	43877	12239	Acceptable

The information given above is believed to be correct, however, both the information and product are offered without warranty for any specific use, other than that specified in the current Oxoid Manual.



Fay Curd
Quality Assurance Manager