

## Certificate of Analysis

**Product Name:** Streptex Latex Gp C 50TST/KT  
**Product Code:** R30950801  
**Lot Number:** 6221678  
**Expiration Date:** 09.03.2026  
**Date of Manufacture:** 06.05.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75WZL53W	ZL53 Strep C Latex Filled Bulk	6179296	09.03.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Abbreviated Latex Efficacy/Specificity	The Latex must show obvious agglutination with its homologous extract in the 10 and 60 minute extraction procedures	Conforms
	The Latex must not show agglutination with the heterologous extracts although a trace granularity is acceptable	Conforms
Compatability of Latex with Streptex Acid Extraction kit	The latex must show obvious agglutination with its homologous extract	Conforms
	The Latex must not show agglutination with the heterologous extracts nor with the extraction fluid from the uninoculated tube although a trace granularity is acceptable	Conforms



Fay Curd  
 Quality Assurance Manager