

## Certificate of Analysis

**Product Name:** Streptex Latex Gp D 50TST/KT  
**Product Code:** R30950901  
**Lot Number:** 6227223  
**Expiration Date:** 12.08.2026  
**Date of Manufacture:** 09.06.2025

\*Date format is DD.MM.YYYY

Item	Description	Lot/Serial	Expire
W75UZL54S	ZL54 Strep D Latex Filled Bulk	6217497	12.08.2026

Test results carried out according to the Quality Assurance Product Release and Test Specification

Test	Anticipated Result	Actual Result
Abbreviated Latex Efficacy/Specificity	The Latex must show obvious agglutination with its homologous extract in the 10 and 60 minute extraction procedures	Conforms
	The Latex must not show agglutination with the heterologous extracts although a trace granularity is acceptable	Conforms



Fay Curd  
Quality Assurance Manager