

CERTIFICATE OF ANALYSIS

PRODUCT CT0066B

E5 ERYTHROMYCIN

LOT NUMBER 6307650

EXPIRY DATE 2028.09.11

DATE OF

MANUFACTURE 2025.09.12

Delivery/Customer information

Date Printed 2025.09.19

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc)

Tested in accordance with legacy BSAC methodology

Iso-Sensitest Agar

Incubated at 35-37°C in air for 18-20 hours

Zone Size	Limits
(mm)	(mm)
26	22 - 31
25	22 - 29
se blood	
28	26 - 33
26	23 - 29
se blood	
hours	
30	23 - 36
27	25 - 29
se blood and with the add	dition of 20mg/l NAD
hours	
12	9 - 16
19	12 - 23
29	20 - 29
	(mm) 26 25 se blood 28 26 se blood hours 30 27 se blood and with the add hours 12 19

5.29

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed.

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.09.19

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.