

CERTIFICATE OF ANALYSIS

PRODUCT	TV5002E	TRYPTONE SOYA BROTH USP (50x10ML)
LOT NUMBER	6259644	
EXPIRY DATE	2026.04.22	
TEST DATE	2025.04.24	
REPORTING DATE	2025.04.29	

General Characteristics	Results	Specification
Colour	Conforms	Maize yellow
Appearance	Conforms	Transparent
pH	7.4	7.1 -7.5
Packaging / Presentation	Conforms	Label & Print check
Filling volume / weight	Conforms	9,5 - 10,5 g
Cont. check @ 20-25 & 30-35°C for >=120h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Bacillus subtilis ATCC®6633	65	Conforms	Turbid growth
Geobacillus stearothermophilus ATCC®7953	39	Conforms	Turbid growth
Candida albicans ATCC®10231	74	Conforms	Turbid, flocculent growth
Aspergillus brasiliensis ATCC®16404	23	Conforms	White mycelia

Tested in accordance with BP/EP/JP/USP. Clearly visible growth within 3 days for bacteria and within 5 days for fungi.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.