

CERTIFICATE OF ANALYSIS

PRODUCT	TV5013D	BUFFERED PEPTONE WATER
LOT NUMBER	6272845	
EXPIRY DATE	2026.05.19	
TEST DATE	2025.05.21	
REPORTING DATE	2025.05.26	

General Characteristics	Results	Specification
Colour	Conforms	Ivory
Appearance	Conforms	Transparent
pH	7.0	6.8 -7.2
Packaging / Presentation	Conforms	Label & Print check
Filling volume / weight (g)	Conforms	8.82 - 9.18
Cont. check @ 20-25 & 30-35°C for >= 72h	Conforms	Within Limits

Microbiological Performance	Control c.f.u	Test Result	Specification
Staphylococcus aureus ATCC®25923 WDCM 00034	59	62	cfu ± 30% of original count
Escherichia coli ATCC®8739 WDCM 00012	95	88	cfu ± 30% of original count
Listeria monocytogenes ATCC®13932 WDCM 00021	69	75	cfu ± 30% of original count
Salmonella Typhimurium ATCC®14028 WDCM 00031	55	Conforms	Turbid growth
Salmonella Enteritidis ATCC®13076 WDCM 00030	82	Conforms	Turbid growth
Escherichia coli ATCC®8739 WDCM 00012	83	Conforms	Turbid growth

Biological performance tested in accordance with current version of ISO 11133.

Formulation conforms to ISO 6579, ISO 6887 and ISO 11290-2.

The information given is believed to be correct. However, both the information and the product are offered without warranty for any specific application other than that specified. The results reported were obtained at the time of release.

This certificate is produced electronically and valid without a signature

The quality control methods meet requirements of ISO 11133.



The testing laboratory of Oxoid Deutschland GmbH is accredited by the German accreditation body DAKKS according to DIN EN ISO/IEC 17025 for the performance testing of media for microbiology to DIN EN ISO11133 and registered under D-PL-20190-01-00.