

CERTIFICATE OF ANALYSIS

PRODUCT CT0034B
F100 NITROFURANTOIN

LOT NUMBER 6255573

EXPIRY DATE 2028.04.08

DATE OF MANUFACTURE 2025.04.09

Delivery/Customer information

Date Printed

2025.04.28

Delivery No.

Customer

Customer Order Number

Microbiological Performance

Assay (mcg/disc) 105.9

Tested in accordance with current EUCAST methodology

	Zone Size (mm)	Limits (mm)
<i>Escherichia coli</i> ATCC®25922	19	17 - 23
<i>Staphylococcus aureus</i> ATCC®29213	19	17 - 23
<i>Enterococcus faecalis</i> ATCC®29212	21	18 - 24
<i>Streptococcus pneumoniae</i> ATCC®49619	25	25 - 31

The above have been tested in our Product Performance Laboratory using conventional procedures and control ATCC and NCTC cultures, where appropriate, and have met the specified test parameters. Additional challenging strains may also be employed. Quality control ranges are internally derived based on historical data

The accuracy of the information provided in this certificate of analysis / conformity is considered reliable. However, it is important to note that (i) the information is only applicable, and (ii) the product is only approved, for the product's intended use as specified on the label, specification, and website. No warranty is hereby implied. This certificate may only be reproduced in its entirety and cannot be replicated partially or in any altered form.

Lot Accepted. 2025.04.28

This certificate is produced electronically and valid without a signature.

ATCC is a registered trade mark of the American Type Culture Collection.

NCTC and National Collection of Type Cultures are registered trade marks of the Health Protection Agency.

OXOID LIMITED

Wade Road, Basingstoke, Hampshire RG24 8PW. England

www.ThermoFisher.com