

## G & M Procter Ltd. Certificate of Analysis

**PRODUCT** **PO0410B**  
**SAB. DEXTROSE AGAR IRRAD + X3 WRAP**  
**1 PACK OF 10 PLATES**

**LOT NUMBER** 4508893  
**EXPIRY DATE** 2025.09.03  
**PACKING DATE** 2025.05.27  
**TEST DATE** 2025.06.04  
**REPORTING DATE** 2025.06.09

All testing in accordance with internally derived specifications, unless otherwise stated.

Physical Characteristics	Results	Specification	Accredited Method Reference
Appearance	Straw 2	Straw 1, straw 1/2 or straw 2	SOP 178 Appearance and colour
pH (25°C)	5.3	5.1 - 5.6	SOP 53 pH
Fill Volume/Weight	18.7g	18.5 - 20.5g	SOP 74 Fill volume weight check
Sterility @ 22°C & 32°C ± 2°C for 5 days	no growth	no growth	SOP 167 Contamination Check at 22°C & 32°C

### MICROBIOLOGICAL PERFORMANCE

For target organisms the control media must achieve a colony count of 10-100 cfu. The test medium must achieve between 70%-150% of the control medium and show the colonial appearance stated in the specification.

Target Organism	Control c.f.u	Test c.f.u	Colonial Appearance	Colonial Appearance Specification	Accredited Method Reference
Candida albicans ATCC®10231	82	91	Cream, domed cols	Cream, domed cols	SOP 151 Fertility of Specified Target Organism(s)(Agar)
Aspergillus brasiliensis ATCC®16404	97	96	White mycelia, black spores	White mycelia, black spores	SOP 151 Fertility of Specified Target Organism(s)(Agar)

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



*Ian Snowball*

Ian Snowball  
Quality Manager  
G & M Procter Ltd.

Performance tested by the Quality Control Laboratory, G & M Procter Ltd, 4 Auld Bond Road,  
Perth, PH1 3FX, a UKAS accredited testing laboratory NO. 2727

## G & M Procter Ltd. Certificate of Analysis

### EXTRA INFORMATION

All results within this section are reported out with the UKAS ISO 17025 Scope of Accreditation.

Tested in accordance with the methods described in the current United States pharmacopoeias for microbial control of aseptic processing environments.

\*Due to process restriction the packing date may be after the test date.

All of the results reported within the G & M Procter Certificate of Analysis relate only to the sample tested. The results were derived from a representative sample of the batch and were obtained at the time of release. The testing laboratory is not responsible or accredited for the sampling process. All test specifications are defined in the G&M Procter manufacturing and test procedures for this product, which are available on request. The uncertainty of measurement introduced during pH, fill weight and microbiological performance testing has been determined, but not reported on the Certificate.



*Ian Snowball*

Ian Snowball  
Quality Manager  
G & M Procter Ltd.

CERTIFICATE OF ANALYSIS

Delivery/Customer information

Date Printed  
2025.06.09  
Delivery No.

Customer

Customer Order number

The information given is believed to be correct. However both the information and the product are offered without warranty for any specific application other than that specified. The results reported were derived from a representative sample of the batch and were obtained at the time of release.



Ian Snowball  
Quality Manager, G&M Procter Ltd

PHYSICAL CHARACTERISTICS	Results	Specification
* Irradiation Dose	Conforms	8-20 kGy
* Irradiation Certificate No.	Gamma Process Run ID 2183-11860A	N/A
Packaging and Presentation	Conforms	Conforms to specification

Our management system is certified by BSI as being in conformity with ISO 9001:2008, certificate number FM 27644 and ISO 13485:2003, certificate number MD 85850.

G & M Procter Ltd, Thermo Fisher Scientific, Microbiology,  
4 Auld Bond Road, Perth, PH1 3FX